

Supplementary Data

This supplementary data is a part of a paper entitled “A Univariate Optimization Strategy for Pre-concentration of Cobalt(II) in Various Matrixes by a DLLME before Analysis Using FAAS”.

Table S1. Furnace program for digest cobalt(II) analysis

Entry	Temperature (°C)	Time (s)	Heat mode	Flow rate (L/min)
1	90	3	RAMP	0.3
2	130	10	RAMP	0.3
3	130	20	STEP	0.3
4	1000	30	RAMP	0.3
5	1000	20	STEP	0.3
6	2400	5	STEP	0.0
7	2550	3	STEP	0.3
8	2550	2	RAMP	0.3